



IONIC® FFPE COMPLETE PURIFICATION KIT

PART #: 33015

FOR RESEARCH USE ONLY. Not intended for therapeutic or diagnostic purposes in humans or animals.
Product guaranteed 6 months from date of receipt.

The FFPE Complete Purification Kit is intended for use only with the Ionic® Purification System. The kit contains fluidic chips and reagents to purify 48 samples; 8 samples per chip.

The Ionic Purification System and FFPE Complete Purification Kit components are intended for research use only.



Ionic® Fluidic Chip

Contents

Description	Quantity	UGM	Volume	Container	Storage Requirements
Fluidic Chips	12	Chip(s)	N/A	Chip Set Box	Room temperature, in unopened antistatic bag
Lysis Buffer 1	12	Bottle(s)	1.4 mL	-20°C Reagents Box	-20°C
RNase Reagent	1	Bottle(s)	700 µL	-20°C Reagents Box	-20°C
DNase Reagent	1	Bottle(s)	700 µL	-20°C Reagents Box	-20°C
Proteinase K Reagent	1	Bottle(s)	1.5 mL	-20°C Reagents Box	-20°C
Lysis Buffer 2	1	Bottle(s)	7.0 mL	RT Reagents Box	Room temperature
Extraction Buffer	1	Bottle(s)	24.0 mL	RT Reagents Box	Room temperature
Anodic Buffer	1	Bottle(s)	24.0 mL	RT Reagents Box	Room temperature
Separation Buffer	2	Bottle(s)	15.0 mL	RT Reagents Box	Room temperature
Neutralization Buffer	1	Bottle(s)	24.0 mL	RT Reagents Box	Room temperature
Cathodic Buffer	1	Bottle(s)	24.0 mL	RT Reagents Box	Room temperature
Mineral Oil	1	Bottle(s)	20.0 mL	RT Reagents Box	Room temperature

TABLE: Ionic FFPE Complete Purification Kit contents

Protocol

Please download the latest protocol for the Ionic FFPE Complete Purification Kit from www.purigenbio.com/support/documentation.

Storage and Handling of Fluidic Chips

Purigen fluidic chips are designed for laboratory use only and must be stored and handled properly to achieve the expected performance. Each chip should remain sealed inside a static dissipative bag and only open prior to immediate use. Chips should be handled only while wearing gloves and gripped by the side skirts. Contact of any kind with the chip surface and features (wells, connection points) should be avoided except for the loading or removal of samples and reagents by pipette. Failure to follow this guidance can result in poor performance or run errors.

PURIGEN BIOSYSTEMS, INC.

5700 Stoneridge Drive, Suite 100
Pleasanton, CA 94588
USA

TEL: +1 877 PURIGEN (787-4436)

SALES

EMAIL: sales@purigenbio.com

SUPPORT

EMAIL: support@purigenbio.com

TEL: +1 877 PURIGEN (787-4436)

FOR RESEARCH USE ONLY. Not for use in diagnostic procedures.

© 2021 Purigen Biosystems Inc. All rights reserved.

The Purigen logo, "Nucleic Acid Purification - Pure and Simple," and Ionic are registered trademarks of Purigen Biosystems, Inc., in the U.S.

